



UNITED STATES DEPARTMENT OF COMMERCE
National Institute of Standards and Technology
Gaithersburg, Maryland 20899-

May 10, 2004

Mr. John Harris
Lab/Cor, Inc.
7619 Sixth Avenue, NW
Seattle, WA 98117

NVLAP Lab Code: 101920-0

Dear Mr. Harris:

On November 7, 2003, your laboratory was visited by an assessor representing the National Voluntary Laboratory Accreditation Program (NVLAP). The purpose of the visit was to assess your laboratory's compliance with NVLAP criteria for accreditation in the Airborne Asbestos Fiber Analysis (TEM) program.

I am pleased to inform you that the On-Site Assessment Review, which was completed on May 7, 2004, has found that your laboratory meets the on-site assessment requirements. No further action is required on your part, at this time, with regard to the on-site assessment phase of the evaluation of your laboratory.

If you have any questions, please call Thomas R. Davis at (301) 975-6499, or Hazel M. Richmond at (301) 975-3024.

Sincerely,

Warren R. Merkel, NVLAP Chief
Laboratory Accreditation Program

NIST

Page 1 of 1

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Proficiency Test Report

Lab Id: 11747 LABCOR INC
 7619 SIXTH AVENUE NW
 EPA Lab Code: SEATTLE, WA-98117-4037
 Not on File (206) 781-0155
 Director: MR. JOHN HARRIS

Shipment Date : 07-Sep-2004
 Closing Date : 22-Oct-2004
 Score Date : 09-Nov-2004

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 276 Asbestos in Air, Water, and NonFriable Samples by Electron Microscopy

| Analyte Name | Units | Sample ID | Method | Result | Mean/Target * | Warning Limits | Acceptance Limits | Score |
|--|---------|-----------|-------------------------|---------|---------------|----------------|-------------------|---|
| Sample: Air and Emissions Asbestos in Air by TEM | | | | | | | | |
| Asbestos in Air by TEM EPA Code: N/A | Struct/ | 1728 | 40 CFR APX A No. III | 41.4 | 88.4 | | D.L. - 196 | Satisfactory 28 passed out of 29 reported results. |
| Dominant Asbestos Type EPA Code: N/A | | 1728 | 40 CFR APX A No. III | Act | Act | | | Satisfactory 16 passed out of 29 reported results. |
| Dominant Asbestos Type EPA Code: N/A | | 1987 | 40 CFR APX A No. III | Amosite | Amosite | | | Satisfactory 27 passed out of 29 reported results. |
| Asbestos in Air by TEM EPA Code: N/A | Struct/ | 1987 | 40 CFR APX A No. III | 10741 | 8040 | | 660 - 15400 | Satisfactory 29 passed out of 29 reported results. |
| Sample: Solid and Hazardous Waste Asbestos in Non-Friable Material | | | | | | | | |
| Percent Asbestos in Residue EPA Code: N/A | % | 1785 | ITEM 198.4 OF MANUAL | 60 | 47.7 | | D.L. - 98.7 | Satisfactory 29 passed out of 29 reported results. |
| Percent Residue EPA Code: N/A | % | 1785 | ITEM 198.4 OF MANUAL | 54.51 | 63.3 | | 28.8 - 97.8 | Satisfactory 28 passed out of 29 reported results. |
| Percent Residue EPA Code: N/A | % | 5706 | ITEM 198.4 OF MANUAL | 17.71 | 18.0 | | 13.2 - 22.8 | Satisfactory 28 passed out of 29 reported results. |
| Percent Asbestos in Residue EPA Code: N/A | % | 5706 | ITEM 198.4 OF MANUAL | 55 | 38.7 | | D.L. - 93.9 | Satisfactory 29 passed out of 29 reported results. |
| Sample: Potable Water Asbestos in Water by TEM | | | | | | | | |
| Asbestos in Water by TEM ** EPA Code: 0253 | MF/L | 4233 | EPA 100.2 | 8.179 | 3.00 * | | D.L. - 19.8 | Satisfactory 22 passed out of 23 reported results. |